



SNS COLLEGE OF TECHNOLOGY

(AN AUTONOMOUS INSTITUTION)

Approved by AICTE & Affiliated to Anna University
Accredited by NBA & Accredited by NAAC with 'A+' Grade,
Recognized by UGC saravanampatti (post), Coimbatore-641035.



Department of Biomedical Engineering

Vision Tit 2

Vision Title 3

Course Name: 19BMB203 Medical Instrumentation

II Year : IV Semester

Unit III – SKELETAL MUSCULAR EQUIPMENT

Topic : ELECTROMYOGRAPHY

19BMB203/Medical Instrumentation/Unit 3 /Mrs.J.Jareena /AP/BME



Electromyography (EMG)

- is a technique for evaluating & recording the electrical activity produced by skeletal muscles. Vision Tit 2
- is performed using an instrument called an electromyography, to produce a record called an electromyogram Vision Title 3
- A **resting muscle does not show recordable electrical potential** but with **increase force of contraction**, amplitude of potential increases
- An electromyography detects electrical potential generated by muscle cells when these cells are electrically or neurologically activated



Equipment

- **Electrodes**
 - Surface electrodes
 - Needle electrodes
- **A high gain amplifier**
- **Connected to an oscilloscope**
 - Oscilloscope traces may be photographed or stored on magnetic tape
- EMG signals may be fed to an audio unit for an on the spot feel of the signals
- EMG is best done in especially constructed shield room to prevent interference

Vision Tit 2


Vision Title 3



Needle Electrode



Vision Title 3



Surface Electrode

In general, three different parameters can be determined by electromyography:


- **Timing** – when is a muscle activated?
- **Force** – How strongly does a muscle contract?
- **Fatigue** – Can a muscle call up its full power?




Pattern of EMG	Recorded	Findings
Resting activity	Muscle relaxed & needle not moving	No activity
Insertion activity	Needle is moved to various sampling spots within insertion tract	Brief action potentials
Motor unit potential	Needle is not moved while patient makes slight contraction	A few motor unit action potentials, biphasic or triphasic, short duration
Recruitment	Subject makes progressively stronger muscle contraction until reaching	Increase number of functioning movements until the baseline is obscured



RESTING EMG




INSERTION ACTIVITY




Vision Tit 2

Vision Title 3

Motor Unit Potential (MUP)




RECRUITMENT PHENOMENON





Motor unit potential (MUP)

- The sum of the action potentials produced in the muscle fibers stimulated by single motor neuron
- Characterized by its duration, number of phases, amplitude, & rate of rise of first component





Feature of MUP

Duration	<ul style="list-style-type: none">• measured from the initial take-off to the point of return to the baseline• 5-15 ms
Phases	<ul style="list-style-type: none">• portion of the MUP between the departure & the return to the baseline• triphasic (positive, negative, positive)• Polyphasic-MUP with more than four phase (5-15%)
Amplitude	<ul style="list-style-type: none">• measured from maximum peak of negative phase to maximum peak of the positive phase• 0.5mV to 2mV
Rise time	<ul style="list-style-type: none">• duration from the initial positive to subsequent negative peak• normal \downarrow 500 μs




Factors that effect MUP

- Technical factors
 - Type of needle electrode
 - Characteristics of recording surface
 - Electrical characteristics of cable
 - Preamplifier & amplifier^{Vision Tit 2}
 - Method of recording
- Physiological factors
 - Age of the patient
 - Muscle examined
 - Temperature



Analysis of EMG Waveform


- **Qualitative** – visual inspection (Size, shape and morphology of EMG signal)
- **Quantitative** – Quantitative information about EmG signal
 - Amplitude of signal
 - Frequency response of the EMG
 - Time Duration
 - Power spectrum





1) The initial control signal to contract a muscle arrives from the brain in the spinal cord. 2) Signal gets transmitted to several motor units. 3) The signal arrives at the muscle and is transformed to a MUAP. The muscle contracts and produces a muscle force. 4) Superposition of individual MUAPs and their sum as the resulting EMG signal. The sum of MUAPs determines the total muscle force.





A Normal muscle



B Denervated muscle



C Myopathy



Vision Title 3



Applications of EMG

Medical Research

- Orthopedic
- Surgery
- Functional Neurology
- Gait & Posture Analysis

Vision Tit 2

Rehabilitation

- Post surgery/accident
- Neurological Rehabilitation
- Physical Therapy
- Active Training Therapy

Vision Tit 3

Ergonomics

- Analysis of demand
- Risk Prevention
- Ergonomics Design
- Product Certification

Sports Science

- Biomechanics
- Movement Analysis
- Athletes Strength Training
- Sports Rehabilitation