



SNS COLLEGE OF TECHNOLOGY

(An Autonomous Institution)

Coimbatore-641035.



ION SELECTIVE ELECTRODE (ISE)

These are the electrodes which have the ability to respond only to a particular ion and develop potential, ignoring the other ions in a mixture.

The potential developed by an ion selective electrode depends only on the concentration of particular ions.

GLASS ELECTRODE

Principle

It has been found experimentally by F.Friitz that when two solutions of different pH values are separated by a thin glass membrane, there develops a difference of potential between the two surfaces of the membrane. The potential difference developed is proportional to the difference in pH value.

The glass membrane functions as an ion exchange resin. Equilibrium is set up between the Na^+ ions of glass and H^+ ions in solution. This forms the basis of glass electrode.


CONSTRUCTION

Glass electrode consists of a thin walled glass bulb. The glass is of a special type of relatively low melting point and high electrical conductivity. The glass bulb may contain AgCl coated silver electrode or a platinum wire in 0.1M HCl.

Hence, the glass electrode may be represented as



(or)



Hydrochloric acid present in the glass bulb furnishes a constant H^+ ion concentration.

Thus, glass electrode is silver –silver chloride electrode, reversible with respect to chloride ion.

Glass electrode is usually employed as a internal reference electrode. The electrode potential varies with H^+ ion concentration and is given by $E_G = E^{\circ}G + 0.0591\text{pH V}$

Hence glass electrode is mostly used to determine the pH of solutions, especially coloured solutions containing oxidising or reducing agent. Usually calomel electrode is used as the second electrode.

Determination of a pH of a solution using glass electrode



SNS COLLEGE OF TECHNOLOGY

(An Autonomous Institution)

Coimbatore-641035.



The glass electrode is placed in the solution under test and is coupled with saturated calomel electrode. The emf of the cell is measured, from which, pH of the solution is calculated.

$$E_{\text{cell}} = E_{\text{right}} - E_{\text{left}}$$

$$E_{\text{cell}} = 0.2422\text{V} - [E^{\circ}_{\text{glass}} + 0.0591 \text{ V pH}]$$

$$E_{\text{cell}} = 0.2422\text{V} - E^{\circ}\text{G} - 0.0591 \text{ V pH}$$

$$\text{pH} = \frac{0.2422\text{V} - E^{\circ}\text{G} - E_{\text{cell}}}{0.0591}$$

ADVANTAGES OF GLASS ELECTRODE

- It can be easily constructed and readily used
- The results are accurate.
- It is not easily poisoned.
- Equilibrium is rapidly achieved.

LIMITATIONS

Since the resistance of glass membrane is quite high, special electronic potentiometers are employed for measurement.

The glass electrode can be used in the pH range of 0 to 10. Above pH 12 (high alkalinity), cations of the solution affect the glass and make the electrode useless.

APPLICATIONS OF ISE

- ISEs are used to determine the concentrations of cations like H^+ , Na^+ , K^+ , Ag^+ , and Li^+
- ISEs are used for the determination of hardness (Ca^{2+} Mg^{2+} ions)
- Concentrations of anions like NO_3^- , CN^- , S^{2-} and halides can also be determined.
- ISEs are used in the determination of concentration of a gas by using gas sensing electrodes.
- pH of the solution can be measured by using gas sensing electrodes.



SNS COLLEGE OF TECHNOLOGY
(An Autonomous Institution)
Coimbatore-641035.

